

Department of Chemistry
Deshbandhu College, University of Delhi,
Kalkaji, New Delhi, 110019

Date: 01.11.2017

Sealed quotations are invited from authorized dealer addressed to 'The Principal, Deshbandhu College, University of Delhi, Kalkaji, New Delhi, 110019 for the purchase of weighing balance for research purpose. The quotations must reach latest by 15th November 2017.

Technical specification for Weighing Balance

Weighing Capacity	:	0 - 220 g
Readability	:	0.0001 g
Pan Size	:	ø 90 mm
Repeatability	:	$\leq \pm 0.0001$ g
Linearity	:	$\leq \pm 0.0002$ g
Response time	:	2.5 (avg., sec.)
Calibration/Adjustment	:	Internal Calibration with Auto Adjustment & External Calibration with Auto Adjustment
Advantage features	:	Monolithic Technology based weigh cell for tolerating the fluctuation in Accuracy with variation in temperature at root level and Tare keys should be both sides.

Terms and Conditions

1. The quotation should contain the complete technical brochure with model number.
2. Please quote for two years warranty with service support in Delhi.
3. Compliance statement to the item specified above should be provided.
4. The college reserves the right to accept or reject any or all the quotations without assigning reasons thereof.
5. The quotation should be valid for 120 days from the date of submission.
6. The price should include all taxes including GST, delivery and installation price etc.
7. Payment will be made after successful installation and satisfactory performance of the equipment.



Dr. Umesh Kumar

Principal Investigator

Project Code -30-311/2016(BSR)



Dr. Ajay K. Arora

Principal (OSD)